

K-ASSAY®

KL-6 Control (Ver.2)

Lot A123, Exp. 2026-02-04

CAT. NO. K352C-2M

INTENDED USE

The **K-ASSAY®** KL-6 Control (Ver.2) is intended for use as a consistent test sample of known concentration for monitoring the performance of the **K-ASSAY®** KL-6 (Krebs von den Lungen-6) (Ver.2) immunoturbidimetric assay. FOR RESEARCH USE ONLY IN THE U.S. NOT FOR USE IN DIAGNOSTIC PROCEDURES IN THE U.S.

SUMMARY

The use of quality control materials to objectively monitor the precision of procedures has been well established. The **K-ASSAY®** KL-6 Control (Ver.2) is provided at two levels.

SET COMPOSITION

Level 1, 2 KL-6 Control (Ver.2) 1 x 2 mL each level
Human KL-6 in 150mM NaCl

WARNINGS AND PRECAUTIONS

For Research Use Only in the U.S. Not for use in diagnostic procedures in the U.S.

Not to be used internally in humans or animals. Normal precautions exercised in handling laboratory reagents should be followed.

Do not mix or use controls from one test kit with those from a different lot number.

Do not use controls past their expiration date stated on each container label.

Do not pipette by mouth. Avoid ingestion and contact with skin.

The controls in this kit do not contain human blood components. However, no assay can be completely free of risk from infection. As such, the material should be handled as potentially infectious in accordance with good laboratory practices and appropriate control. See the National Institute of Health Manual, "Biosafety in Microbiology and Biomedical Laboratories," 2nd ed., 1988.

Controls in this kit contain < 0.1% w/v sodium azide as a preservative. Sodium azide may form explosive compounds in metal drain lines. When disposing of reagents through plumbing fixtures, flush with copious amounts of water. For further information, refer to "Decontamination of Laboratory Sink Drains to Remove Azide Salts," in the Manual Guide-Safety Management No. CDC-22 issued by the Centers for Disease Control, Atlanta, GA.

CONTROL PREPARATION

The controls are ready to use. They do not require reconstitution.

STORAGE AND HANDLING

Store controls at 2-8°C. Return all controls to 2-8°C promptly after use. Unopened controls can be used for up to 1 year from the date of manufacture, as indicated by the expiration date on the package and bottle labels.

CONTROL STABILITY

Opened bottles of controls can be used for 1 month if stored at 2-8°C.

Discard controls if they become contaminated. Evidence of cloudiness or particulate material in solution is cause to discard.

PROCEDURE

Materials Supplied

Level 1 1 x 2 mL
Level 2 1 x 2 mL

Materials Required But Not Supplied

K-ASSAY® KL-6 (Ver.2) immunoturbidimetric assay

K-ASSAY® KL-6 Calibrator (Ver.2)

Two Reagent Clinical Chemistry Analyzer Capable of:
Accurate absorbance readings at approx. 570 nm
Accurately dispensing the required volumes
Maintaining 37°C

Assay Procedure

NOTE: Allow all reagents and specimens to come to room temperature.

The **K-ASSAY®** KL-6 Control (Ver.2) is assayed using the same procedure as the samples run in the test procedure. See package insert from the **K-ASSAY®** KL-6 (Ver.2) immunoturbidimetric assay kit.

LIMITATIONS OF PROCEDURE

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, and good laboratory technique. Erroneous results can occur from improper storage, reconstitution inaccuracies, and technical errors associated with assay procedures.

This product is intended for use as an assayed control for quantitative assays of KL-6 in human serum and plasma. This product is not intended for use as a calibrator.

EXPECTED VALUES

KAMIYA BIOMEDICAL COMPANY has established the expected values. The assignment of mean values was derived from analysis of vials representative of the entire lot.

Values listed are specific for each lot only. Verify that the lot numbers on the vials of **K-ASSAY**® KL-6 Control (Ver.2) correspond to the lot numbers listed for the Assay Data.

The Expected Range of the Mean is provided to assist the laboratory until it has established its own mean and standard deviation. It is considered good laboratory practice for each laboratory to establish its own mean and standard deviation for its test methods. The indicated Mean and Expected Range of the Mean should serve as a guide in assessing the performance of each test method.











ASSAY DATA

Lot A123

ASSAY	UNIT	LEVEL 1		LEVEL 2	
		MEAN	RANGE	MEAN	RANGE
K-ASSAY ® KL-6 (Ver.2)	U / mL	350	298 – 403	1,000	850 – 1,150

The expected values for the **K-ASSAY**® KL-6 Control (Ver.2) are continually being revised through ongoing quality assurance. Please refer to the package insert included with each control set for the most appropriate control values. Recovered values may be method dependent. The variations which can occur over time and between laboratories may be attributed to differences in laboratory technique, instrumentation, reagent lot, method modifications, and other systematic errors including random errors.

LABELING SYMBOLS

	Catalog Number
	Expiration or "Use By" Date
	Lot Number
	Control
	Consult Package Insert for Instructions for Use
	CE Mark Registered
	Potential Human Biohazard
	Temperature Limitation. Store between 2 and 8 degrees C
	Manufacturer
	Authorized Representative in the European Community

EU AUTHORIZED REPRESENTATIVE



Advena Ltd.
Tower Business Centre, 2nd Flr.,
Tower Street, Swatar, BKR 4013 Malta

ORDERING / PRICING / TECHNICAL INFORMATION



KAMIYA BIOMEDICAL COMPANY
12779 Gateway Drive
Seattle, WA 98168 USA
TEL: (206) 575-8068 / (800) 526-4925
FAX: (206) 575-8094

Printed February 2025